

Certificate of Analysis

Mycoplasma Gallisepticum Antibody Test Kit

Product Code: CK114

Lot No: **FS9488**

Expiry Date: **6th November 2024**

	Kit Component	Lot No		Kit Component	Lot No
1	Coated Plates	CH7684	6	Sample Diluent Reagent	CH8049
2	Conjugate Reagent	CH8230	7	Wash Buffer Powder	CH6320
3	Substrate Tablets	CH8147	8	Negative Control	CH6859
4	Substrate Reagent	CH8063	9	Positive Control	CH8231
5	Stop Solution	CH8054			

Quality Control Results

Mean Negative Control

Optical Density = 0.092

Mean Positive Control

Optical Density = 0.635

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF18)

Antibody titre = 1020
(range: 800 – 1500)

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF19)

Antibody titre = 2424
(range: 2000 – 5000)

Date of test: 6th November 2023

The above result was generated by performing the test according to the package insert.

The above information is correct and forms part of the BioChek Quality Control for Mg ELISA kit product code CK114.



Released by QC – 9th November 2023



Jaroslaw Wnuk
Quality Control Manager