

Certificate of Analysis

Egg Drop Syndrome Virus Antibody Test Kit

Product Code: CK112

Lot No: FS9534

Expiry Date: 6th December 2024

	Kit Component	Lot No		Kit Component	Lot No
1	Coated Plates	CH7317	6	Sample Diluent Reagent	CH8195
2	Conjugate Reagent	CH8366	7	Wash Buffer Powder	CH6318
3	Substrate Tablets	CH8246	8	Negative Control	CH6861
4	Substrate Reagent	CH7839	9	Positive Control	CH8367
5	Stop Solution	CH8053			

Quality Control Results

Mean Negative Control

Optical Density = 0.135

Mean Positive Control

Optical Density = 0.714

Pre-diluted Reference Control
(Product Code CR100, Lot no. RF18)

Antibody titre = 876
(range: 650 – 2000)

Date of test: 6th December 2023

The above result was generated by performing the test according to the package insert.

The above information is correct and forms part of the BioChek Quality Control for EDS ELISA kit product code CK112.



Released by QC – 8th December 2023



Andrew Russell
Quality Control Supervisor